

PT. B.D.SHARMA PGIMS/UHS ROHTAK
e-TENDER NOTICE NO.UHSR/PURCHASE/COVID-19/2020/03 DT. 24.04.2020
(Due on 01.05.2020 upto 4.00 PM and to be opened on 01.05.2020 at 05.00 PM)
e-Tender Portal: www.etenders.hry.nic.in, Institute website: www.uhsr.ac.in

CORRIGENDUM

E - Tender Group 01

Sr. No.	Specifications in original E- Tender	Specifications now after corrigendum
9. (d)	Maximum inspiratory Peak Flow 200 l/min or more	Maximum inspiratory Peak Flow 200 l/min or more. Minor deviations in maximum inspiratory peak flow can be accepted.
15.	Alarm History of -> 1000 alarms.	Alarm History of -> 500 alarms.
18.	Displayed Trends Values for 48-72 hours at least for above parameters.	Displayed Trends Values for atleast 48 hours at least for above parameters.

Note: F.O.R. PGIMS, Rohtak, other Government Medical Colleges and Govt. aided Medical Colleges in Haryana.

Sd/-
Professor I/c Purchase